

Routine Name	Run #	Date & Time
--------------	-------	-------------

gage	1	2025-03-24
------	---	------------

Feature	Unit	Nominal	Actual	Tolerances	Deviation	Exceeded
---------	------	---------	--------	------------	-----------	----------

Step 1 - Item_1 distancia

X Location	mm	-50.0000	-49.9952	+0.0000	-0.0000	-0.0048
------------	----	----------	----------	---------	---------	---------

Step 2 - Item_2 distancia

■ X Location	mm	-103.0000	-103.0508	+0.1000	-0.0000	+0.0508	+++
--------------	----	-----------	-----------	---------	---------	---------	-----

Step 3 - Item_3 distancia

■ X Location	mm	-125.5000	-125.5491	+0.1000	-0.0000	+0.0491	++
--------------	----	-----------	-----------	---------	---------	---------	----

