




Resultados de la calibración  
(Calibration results)

| Cota  | Valor nominal<br>(Nominal value) | Valor mínimo<br>(Minimal value) | Valor máximo<br>(Maximum value) | Promedio<br>(Average) | Desviación estándar<br>(Standard deviation) | Incertidumbre de medida<br>(Measurement uncertainty) |
|-------|----------------------------------|---------------------------------|---------------------------------|-----------------------|---------------------------------------------|------------------------------------------------------|
|       | mm                               | mm                              | mm                              | mm                    | mm                                          | mm                                                   |
| d     | 0.560 0                          | 0.512 0                         | 0.525 0                         | 0.519 7               | 0.003 1                                     | 0.003 1                                              |
| W (x) | 1.000 0                          | 1.008 0                         | 1.038 0                         | 1.023 6               | 0.006 8                                     | 0.003 2                                              |
| W (y) | 1.000 0                          | 0.972 0                         | 1.022 0                         | 0.995 7               | 0.013 1                                     | 0.003 6                                              |



Condiciones del instrumento:  
(Instrument conditions)

Sin comentarios

Requerimientos del cliente:  
(Customer requirements)

Sin requerimientos.