

Routine Name				Run #	Date & Time		
gage				1	2025-03-24		
Feature	Unit	Nominal	Actual	Tolerances		Deviation	Exceeded
Step 1 - Item_1 distancia							
X Location	mm	-50.0000	-49.9952	+0.0000	-0.0000	-0.0048	
Step 2 - Item_2 distancia							
■ X Location	mm	-103.0000	-103.0508	+0.1000	-0.0000	+0.0508	+++
Step 3 - Item_3 distancia							
■ X Location	mm	-125.5000	-125.5491	+0.1000	-0.0000	+0.0491	++

